

**MINUTES OF MEETING OF GRIEVANCES COMMITTEE FOR THE REDRESSAL OF GRIEVANCES OF FIRM AGAINST THE TECHNICAL SCRUTINY OF TENDER FOR THE PLACEMENT OF (08 NOS. BRAND NEW HAEMATOLOGY ANALYZER (03 PARTS) CLOSE SYSTEM FOR THE YEAR 2023-24 ON REAGENT RENTAL BASIS FOR THE YEAR 2023-24 MEETING HELD ON 24-11-2023 AT 10:00 AM IN THE MEETING ROOM OF PURCHASE CELL OMC BAHAWALPUR**

The Meeting started with recitation from Holy Quran. The proceedings were started as per agenda. And decision was made as under:-  
**Agenda of the meeting as under:-**

Sr. #	Name of Firm	Decision of TAC	Grievances Of Firm	Decision
1.	M/S Martin Dow Lahore	Eligible	The firm submitted the grievance Pleaded against M/S A.M Sale & Service, that the quoted model of equipment can analyze only 19 parameter, while the minimum required specification more 20 parameter	The Committee scrutinized the record and discussed the matter in detail with end user. The end user was of opinion that most common reporting parameter for complete blood count were available; the rest of parameter were for research purpose and not included in clinical reporting. In the light of opinion of end user and for better competition. Committee upheld the decision of Technically advisory committee. Furthermore, The committee advised M/S A.M Sale & Service to submit an affidavit required provision of Hematology Analyzer 3parts which can analyze more than 20 parameter at a time.