

**BRAND NEW 02 NOS. FULLY AUTOMATED COAGULATION ANALYZER ON REAGENT RENTAL BASIS FOR THE YEAR 2023-24****TECHNICAL EVALUATION COMMITTEE REPORT**

<b>Name of Document</b>	<b>M/S BioMadics</b>	<b>M/S A.M Sale &amp; Service</b>	<b>M/S Martin Dow Pvt. Ltd</b>	<b>M/S Popular International</b>
<b>Tender Purchase Receipt Attached</b>	Yes	Yes	Yes	Yes
<b>NIC attested copy attached</b>	Yes	Yes	Yes	Yes
<b>NTN attested copy attached</b>	Yes	Yes	Yes	Yes
<b>STN attested copy attached</b>	Yes	Yes	Yes	Yes
<b>One Year Business History</b>	<b>NO. Not Relevant History</b>	Yes	<b>NO. Not Relevant History</b>	<b>NO. Not Relevant History</b>  <b>Poor Feed Back from end user after sale &amp; service</b>
<b>Authority Letter / distribution certificate of manufacture / Importer</b>	Yes	Yes	Yes	Yes
<b>Establishment License/ License to import</b>	Yes	Yes	Yes	Yes
<b>Name of Manufacturer of instrument</b>	Diagon	STAGO	STAGO	Human Diagnostic
<b>Model Name and Model No.</b>	Coag M	STA Compact	STA Compact Max 3	Huma Clot Pro
<b>Country of origin</b>	Hungry	France	France	Germany
<b>Country of manufacturing of instrument</b>	Hungry	France	France	Germany
<b>Quality Certificate</b>	CE Expired ISO-2024	Yes	<b>NO.</b>	Yes

**BRAND NEW 02 NOS. FULLY AUTOMATED COAGULATION ANALYZER ON REAGENT RENTAL BASIS FOR THE YEAR 2023-24**

**TECHNICAL EVALUATION COMMITTEE REPORT**

<b>Name of Document</b>	<b>M/S BioMadics</b>	<b>M/S A.M Sale &amp; Service</b>	<b>M/S Martin Dow Pvt. Ltd</b>	<b>M/S Popular International</b>
<b>Affidavit regarding non black listing of the firm</b>	Yes	Yes	Yes	Yes
<b>Affidavit regarding Price Reasonability Certificate of the firm</b>	Yes	Yes	Yes	Yes
<b>Technically Eligibility of the firm</b>	Not Eligible	<b>Eligible</b>	Not Eligible	Not Eligible

**BRAND NEW 02 NOS. FULLY AUTOMATED COAGULATION ANALYZER ON REAGENT RENTAL BASIS FOR THE YEAR 2023-24**  
**TECHNICAL EVALUATION COMMITTEE REPORT**

	<b>M/S BioMadies</b>	<b>M/S A.M Sale &amp; Service</b>	<b>M/S Martin Dow Pvt. Ltd</b>	<b>M/S Popular International</b>
<p><b>Detailed Requirement:</b></p> <p>Throughput: 50 Tests /hr</p> <p>Random access system fully automated</p> <p>RS 232 or USB Interface</p> <p>220V 50 Hz, AC</p> <p>Up to 8 Reagent Positions on board</p> <p>Up to 80 Samples Positions on board for tubes and cups</p> <p>On Board storage capacity of cuvettes with auto refill without stopping the operation</p> <p>Digital display for results</p> <p>Country of origin: Europe/ USA Japan</p> <p>CE Marked/ FDA/ IVD Certified</p>	YES	YES	YES	YES
<p><b><u>Methodology</u></b></p> <p>Should be able to measured by for clotting. Chromgenie &amp;immunologic assay.</p>	YES	YES	YES	YES

**BRAND NEW 02 NOS. FULLY AUTOMATED COAGULATION ANALYZER ON REAGENT RENTAL BASIS FOR THE YEAR 2023-24**

**TECHNICAL EVALUATION COMMITTEE REPORT**

<p><b><u>Mandatory</u></b></p> <p>Calibrators (Free of cost)</p> <p>Replacement of Parts (Free of cost)</p> <p>Controls Level I,II,III services &amp; maintenance (Free of cost)</p> <p>Barcode identification</p> <p>Visit of engineer (Free of cost)</p>	<p align="center">YES</p>	<p align="center">YES</p>	<p align="center">YES</p>	<p align="center">YES</p>
<p><b><u>Accessories:</u></b></p> <p>Complete with standard and operation accessories;</p> <ol style="list-style-type: none"> <li>1. Built-in or External Laser Printer</li> <li>2. Compatible Computer, LCD Monitor</li> <li>3. Imported Compatible Sine wave UPS for back up of up to 30 minutes (Emerson, Liebert, Chloride, MGE, APC or Equivalent)</li> <li>4. Operating Manual with a Soft Copy</li> <li>5. Service Manual with a Soft Copy</li> <li>6. One Backup Model with same specification</li> <li>7. Microsoft built system</li> <li>8. Centrifuge Machine Axial head 12 tubing</li> </ol>	<p align="center">YES</p>	<p align="center">YES</p>	<p align="center">YES</p>	<p align="center">YES</p>

**BRAND NEW 02 NOS. FULLY AUTOMATED COAGULATION ANALYZER ON REAGENT RENTAL BASIS FOR THE YEAR 2023-24**

**TECHNICAL EVALUATION COMMITTEE REPORT**

9. AC for temperature maintenance of room at (20-25C°)				
	<b>M/S BioMadies</b>	<b>M/S A.M Sale &amp; Service</b>	<b>M/S Martin Dow Pvt. Ltd</b>	<b>M/S Popular International</b>
<b>PT</b>	Yes	Yes	Yes	Yes
<b>APTT</b>	Yes	Yes	Yes	Yes
<b>DECISION OF TECHNICAL ADVISORY COMMITTEE</b>	Not Eligible	<b>Eligible</b>	Not Eligible	Not Eligible